



PINNACLE — ANALYTICS —

Potency Results

Sample Name: *Blue Dream - Delta 8*
Client: Just Kana
Client Batch ID:

Pinnacle-Analytics.com
3549 Lear Way, Suite 101
Medford OR 97504
P:(541)300-8217

Sample ID: rC-H-519-E2013

Date Sampled: 9/19/2024

Matrix: Flower

Date Reported: 9/24/2024

Prep Analyst: Jeff A.

Client License: N/A

Analysis Method: 0668534+1 H4 5-24-2024 #1.lcm

3433 Wilshire Blvd. Suite 2765

Sampling Method: N/A

Los Angeles CA 90010

Reference Method: JCB 2009: HPLC/DAD

For R&D Purposes Only


Analysis Batch: 9-23-2024 H4 390, 519 Flower

Total THC (THCA*0.877+d9-THC)	1.47%
Total CBD (CBDA*0.877+CBD)	14.1%
Moisture Content	9.32%

Cannabinoid	% Weight	mg/g
CBDVA	<LOQ	<LOQ
CBDV	<LOQ	<LOQ
CBDA*	10.0	100.0
CBGA	0.718	7.18
CBG	0.243	2.43
CBD*	5.3	53.0
THCV	<LOQ	<LOQ
CBN	0.127	1.27
d9-THC*	1.37	13.7
d8-THC*	8.42	84.2
CBC	0.354	3.54
THCA*	0.114	1.14
Total Cannabinoids	26.65	266.0

*ORELAP Accredited Analyte

Limit Of Quantitation: 0.1%, analyte not measured



CBDA*	CBD*	d8-THC*
CBGA	CBN	CBC
CBG	d9-THC*	THCA*



These test results may not be altered or reproduced except in full without the permission of Pinnacle Analytics. These results were generated following the Oregon Administrative Rules and in accordance with the NELAP Institute under ORELAP License #4152
Report generated by Routine_Potency_Rev13_8-1-2023

Kris Ford, PhD
Lab Director



PINNACLE ANALYTICS

Quality Control Results

Analyst: Jeff A.

Analysis Batch: 9-23-2024 H4 390, 519 Flower

Pinnacle-Analytics.com
3549 Lear Way, Suite 101
Medford OR 97504
P:(541)300-8217

	Duplicate RPD		LCS % Recovery		Method Blank	
	H-0-E4165-b	Limit	C-FL-092324	Limits	C-FB-092324	Limit
CBDA	3.26%	10%	99.1%	90-110%	<LOQ/2	LOQ/2
CBD	2.44%	10%	97.4%	90-110%	<LOQ/2	LOQ/2
d9-THC	<LOQ%	30%	97.4%	90-110%	<LOQ/2	LOQ/2
d8-THC	<LOQ%	30%	97.9%	90-110%	<LOQ/2	LOQ/2
THCA	<LOQ%	30%	99.0%	90-110%	<LOQ/2	LOQ/2

RPD: Relative Percent Difference between unknown sample and its duplicate

LCS: Laboratory Control Sample with known concentration

Case Comments: There were no divergences from ordinary Quality Control procedures or SOPs.



These test results may not be altered or reproduced except in full without the permission of Pinnacle Analytics. These results were generated following the Oregon Administrative Rules and in accordance with the NELAP Institute under ORELAP License #4152
Report generated by Routine_Potency_Rev13_8-1-2023

Kris Ford, PhD
Lab Director